

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
 Data File : VV035403.D
 Acq On : 25 Apr 2024 14:50
 Operator : SY/MD
 Sample : VV0425WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK276

Quant Time: Apr 26 01:37:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:37:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	380422	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	371456	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	157059	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	189563	51.597	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 103.200%		
7) Chloroethane-d5	1.519	69	149411	46.822	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 93.640%		
11) 1,1-Dichloroethene-d2	2.053	65	77199	47.936	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 95.880%		
21) 2-Butanone-d5	3.783	46	179555	82.405	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 82.400%		
24) Chloroform-d	4.243	84	311594	50.444	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.880%		
26) 1,2-Dichloroethane-d4	4.934	65	193235	49.272	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 98.540%		
32) Benzene-d6	4.953	84	622649	49.902	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 99.800%		
36) 1,2-Dichloropropane-d6	5.982	67	196634	49.628	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 99.260%		
41) Toluene-d8	7.239	98	531711	47.318	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 94.640%		
43) trans-1,3-Dichloroprop...	7.551	79	81705	42.418	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 84.840%		
47) 2-Hexanone-d5	8.024	63	93036	62.105	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 62.110%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	258774	47.427	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 94.860%		
66) 1,2-Dichlorobenzene-d4	11.557	152	186258	52.819	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 105.640%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
Data File : VV035403.D
Acq On : 25 Apr 2024 14:50
Operator : SY/MD
Sample : VV0425WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK276

Quant Time: Apr 26 01:37:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:37:05 2024
Response via : Initial Calibration

