

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042524\
 Data File : VW035404.D
 Acq On : 25 Apr 2024 15:12
 Operator : SY/MD
 Sample : P2158-14
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Apr 26 01:37:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:37:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	366903	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	367995	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	154476	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	178297	50.318	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.640%		
7) Chloroethane-d5	1.523	69	72369	23.514	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 47.020%#		
11) 1,1-Dichloroethene-d2	2.053	65	72245	46.512	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 93.020%		
21) 2-Butanone-d5	3.796	46	188089	89.502	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 89.500%		
24) Chloroform-d	4.246	84	304647	51.136	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 102.280%		
26) 1,2-Dichloroethane-d4	4.941	65	191287	50.573	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 101.140%		
32) Benzene-d6	4.957	84	607234	49.125	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 98.240%		
36) 1,2-Dichloropropane-d6	5.986	67	193584	49.317	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 98.640%		
41) Toluene-d8	7.243	98	520246	46.733	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 93.460%		
43) trans-1,3-Dichloroprop...	7.555	79	82293	43.125	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 86.240%		
47) 2-Hexanone-d5	8.027	63	103974	70.060	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 70.060%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	269702	49.895	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 99.800%		
66) 1,2-Dichlorobenzene-d4	11.558	152	189617	54.671	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 109.340%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
Data File : VV035404.D
Acq On : 25 Apr 2024 15:12
Operator : SY/MD
Sample : P2158-14
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Apr 26 01:37:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:37:05 2024
Response via : Initial Calibration

