

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042524\
 Data File : VW035407.D
 Acq On : 25 Apr 2024 16:29
 Operator : SY/MD
 Sample : P2238-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Apr 26 01:38:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 26 01:37:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	380175	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	376433	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	157020	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	167149	45.525	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.060%
7) Chloroethane-d5	1.529	69	134825	42.278	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.560%
11) 1,1-Dichloroethene-d2	2.053	65	73336	45.566	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.140%
21) 2-Butanone-d5	3.799	46	202291	92.900	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.900%
24) Chloroform-d	4.243	84	313395	50.768	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.540%
26) 1,2-Dichloroethane-d4	4.937	65	195673	49.926	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.860%
32) Benzene-d6	4.957	84	628653	49.717	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.440%
36) 1,2-Dichloropropane-d6	5.985	67	204280	50.876	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.760%
41) Toluene-d8	7.239	98	539240	47.354	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.700%
43) trans-1,3-Dichloroprop...	7.551	79	82728	42.381	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.760%
47) 2-Hexanone-d5	8.030	63	108199	71.272	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.270%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	278640	50.393	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.780%
66) 1,2-Dichlorobenzene-d4	11.558	152	195312	55.401	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042524\
Data File : VV035407.D
Acq On : 25 Apr 2024 16:29
Operator : SY/MD
Sample : P2238-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Apr 26 01:38:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 26 01:37:05 2024
Response via : Initial Calibration

