

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042619\
 Data File : VV010499.D
 Acq On : 25 Apr 2019 18:11
 Operator : SY/MD
 Sample : VSTD0.103
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.103

Quant Time: Apr 26 08:05:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 08:02:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	6172	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5871	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2826	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.32	65	422	0.05	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.09	65	388	0.09	ug/L	0.02
7) Toluene-d8	7.35	98	1130	0.09	ug/L	0.00
9) 1,1,2,2-Tetrachloroethane-	10.25	84	322	0.08	ug/L	0.00
11) 1-2-dichlorobenzene-d4	11.67	152	555	0.10	ug/L	0.00
Target Compounds						
3) Vinyl chloride	1.32	62	462	0.061	ug/L #	40
6) Trichloroethene	5.96	95	474	0.090	ug/L	93
8) 1,2-Dibromoethane	8.40	107	279	0.094	ug/L #	90
12) 1,2,3-Trichloropropane	10.32	75	225	0.077	ug/L	89
13) 1,2-Dibromo-3-chloropropan	12.48	75	72	0.127	ug/L #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042619\
Data File : VV010499.D
Acq On : 25 Apr 2019 18:11
Operator : SY/MD
Sample : VSTD0.103
Misc : 25mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.103

Quant Time: Apr 26 08:05:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 08:02:06 2019
Response via : Initial Calibration

