

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042619\
 Data File : VV010503.D
 Acq On : 25 Apr 2019 20:22
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV07

Quant Time: Apr 26 08:29:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 08:26:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	6240	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5862	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2723	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2419	0.52	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.00%
4) 1,2-Dichloroethane-d4	5.07	65	2211	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
7) Toluene-d8	7.35	98	6438	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
9) 1,1,2,2-Tetrachloroethane-	10.25	84	1820	0.54	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	108.00%
11) 1-2-dichlorobenzene-d4	11.67	152	2784	0.51	ug/L	0.00
Spiked Amount	0.500			Recovery	=	102.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	2401	0.515	ug/L 100
6) Trichloroethene	5.96	95	2566	0.519	ug/L 99
8) 1,2-Dibromoethane	8.40	107	1495	0.526	ug/L 99
12) 1,2,3-Trichloropropane	10.30	75	1285	0.552	ug/L 99
13) 1,2-Dibromo-3-chloropropan	12.46	75	275	0.434	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042619\
Data File : VV010503.D
Acq On : 25 Apr 2019 20:22
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV07

Quant Time: Apr 26 08:29:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 08:26:41 2019
Response via : Initial Calibration

