

Method Path : Z:\V0ASRV\HPCHEM1\MSV0A_V\METHOD\
 Method File : SOMVTR042619SIM.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Apr 26 08:26:41 2019
 Response Via : Initial Calibration

Calibration Files

0.5 =VV010500.D 0.05=VV010498.D 0.1 =VV010499.D
 1.0 =VV010501.D 2 =VV010502.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.415	0.312	0.342	0.386	0.395	0.370	11.36
3) T	Vinyl chloride	0.418	0.330	0.374	0.369	0.376	0.373	8.43
4) S	1,2-Dichloroethane-	0.378	0.257	0.314	0.361	0.373	0.337	15.24

5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.467	0.381	0.404	0.417	0.439	0.422	7.84
7) S	Toluene-d8	1.153	0.800	0.962	1.113	1.185	1.043	15.38
8) T	1,2-Dibromoethane	0.271	0.190	0.238	0.250	0.262	0.242	13.06
9) S	1,1,2,2-Tetrachloro	0.322	0.220	0.274	0.314	0.316	0.289	14.82

10) I	1,4-Dichlorobenzene-d	-----ISTD-----						
11) S	1-2-dichlorobenzene	1.006	1.016	0.982	0.977	0.999	0.996	1.62
12)	1,2,3-Trichloroprop	0.485	0.336	0.398	0.449	0.468	0.427	14.13
13) T	1,2-Dibromo-3-chlor	0.100	0.170	0.127	0.092	0.092	0.116	28.58

(#) = Out of Range