

Data Path : Z:\voasrv\HPCHEM1\MSVOA V\Data\VV042619\
 Data File : VV010503.D
 Acq On : 25 Apr 2019 20:22
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV07

Quant Time: Apr 26 08:29:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 08:26:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
2 S	Vinyl Chloride-d3	0.370	0.388	-4.9	97	0.00
3 T	Vinyl chloride	0.373	0.385	-3.2	96	0.00
4 S	1,2-Dichloroethane-d4	0.337	0.354	-5.0	98	0.00
5 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
6 T	Trichloroethene	0.422	0.438	-3.8	96	0.00
7 S	Toluene-d8	1.043	1.098	-5.3	98	0.00
8 T	1,2-Dibromoethane	0.242	0.255	-5.4	97	0.00
9 S	1,1,2,2-Tetrachloroethane-d	0.289	0.310	-7.3	99	0.00
10 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
11 S	1-2-dichlorobenzene-d4	0.996	1.022	-2.6	101	0.00
12	1,2,3-Trichloropropane	0.427	0.472	-10.5	96	0.00
13 T	1,2-Dibromo-3-chloropropane	0.116	0.101	12.9	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0