

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042619\
 Data File : VV010498.D
 Acq On : 25 Apr 2019 16:53
 Operator : SY/MD
 Sample : VSTD0.0502
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.0502

Quant Time: Apr 26 08:20:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 08:13:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	6309	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5989	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2944	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.32	65	197	0.04	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.09	65	162	0.04	ug/L	0.02
7) Toluene-d8	7.36	98	479	0.04	ug/L	0.01
9) 1,1,2,2-Tetrachloroethane-	10.25	84	132	0.04	ug/L	0.00
11) 1-2-dichlorobenzene-d4	11.67	152	299	0.05	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Vinyl chloride	1.32	62	208	0.044	ug/L #	1
6) Trichloroethene	5.96	95	228	0.045	ug/L	91
8) 1,2-Dibromoethane	8.40	107	114	0.039	ug/L #	86
12) 1,2,3-Trichloropropane	10.32	75	99	0.039	ug/L	92
13) 1,2-Dibromo-3-chloropropan	12.49	75	50	0.073	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042619\
Data File : VV010498.D
Acq On : 25 Apr 2019 16:53
Operator : SY/MD
Sample : VSTD0.0502
Misc : 25mL/MSVOA V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.0502

Quant Time: Apr 26 08:20:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 08:13:33 2019
Response via : Initial Calibration

