

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042619\
 Data File : VV010500.D
 Acq On : 25 Apr 2019 19:03
 Operator : SY/MD
 Sample : VSTD0.504
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.504

Quant Time: Apr 26 08:05:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 08:02:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	5982	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5702	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2751	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.32	65	2484	0.31	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.07	65	2264	0.52	ug/L	0.00
7) Toluene-d8	7.35	98	6573	0.52	ug/L	0.00
9) 1,1,2,2-Tetrachloroethane-	10.25	84	1837	0.50	ug/L	0.00
11) 1-2-dichlorobenzene-d4	11.67	152	2767	0.49	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Vinyl chloride	1.32	62	2503	0.342	ug/L	100
6) Trichloroethene	5.96	95	2662	0.523	ug/L	100
8) 1,2-Dibromoethane	8.40	107	1546	0.536	ug/L	100
12) 1,2,3-Trichloropropane	10.30	75	1334	0.469	ug/L	100
13) 1,2-Dibromo-3-chloropropan	12.46	75	276	0.501	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042619\
Data File : VV010500.D
Acq On : 25 Apr 2019 19:03
Operator : SY/MD
Sample : VSTD0.504
Misc : 25mL/MSVOA V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.504

Quant Time: Apr 26 08:05:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 26 08:02:06 2019
Response via : Initial Calibration

