

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042619\
 Data File : VV010509.D
 Acq On : 26 Apr 2019 14:56
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.509

Quant Time: Apr 27 04:47:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 27 04:44:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.65	114	6294	0.50	ug/L	-0.02
5) Chlorobenzene-d5	8.89	117	6001	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.30	152	2832	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2579	0.55	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.00%
4) 1,2-Dichloroethane-d4	5.07	65	2420	0.57	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	114.00%
7) Toluene-d8	7.35	98	6838	0.55	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	110.00%
9) 1,1,2,2-Tetrachloroethane-	10.25	84	1979	0.57	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.00%
11) 1-2-dichlorobenzene-d4	11.67	152	2919	0.52	ug/L	0.00
Spiked Amount	0.500	Range	80 - 120	Recovery	=	104.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	2391	0.509 ug/L	81
6) Trichloroethene	5.95	95	2658	0.525 ug/L	87
8) 1,2-Dibromoethane	8.40	107	1567	0.539 ug/L	99
12) 1,2,3-Trichloropropane	10.30	75	1382	0.571 ug/L	98
13) 1,2-Dibromo-3-chloropropan	12.46	75	277	0.420 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042619\
Data File : VV010509.D
Acq On : 26 Apr 2019 14:56
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.509

Quant Time: Apr 27 04:47:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 27 04:44:38 2019
Response via : Initial Calibration

