

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042621\
 Data File : VW021150.D
 Acq On : 26 Apr 2021 18:23
 Operator : SY/MD
 Sample : M2165-16ME
 Misc : 5.56g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB3F7ME

Quant Time: Apr 27 04:18:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 27 04:16:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	528513	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.860	117	526387	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.262	152	246812	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	216759	49.91	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	99.82%		
7) Chloroethane-d5	1.571	69	173613	54.34	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	108.68%		
11) 1,1-Dichloroethene-d2	2.092	63	334273	41.73	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	83.46%		
21) 2-Butanone-d5	3.918	46	309903	146.13	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery =	146.13%#		
24) Chloroform-d	4.349	84	581252	75.08	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	150.16%#		
26) 1,2-Dichloroethane-d4	5.034	65	414754	84.27	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	168.54%#		
32) Benzene-d6	5.047	84	1015002	72.40	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	144.80%#		
36) 1,2-Dichloropropane-d6	6.072	67	330974	79.09	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	158.18%#		
41) Toluene-d8	7.317	98	942924	69.50	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	139.00%#		
43) trans-1,3-Dichloroprop...	7.625	79	167501	73.25	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	146.50%#		
47) 2-Hexanone-d5	8.133	63	329816	198.72	ug/L	0.04
Spiked Amount 100.000	Range 45 - 130		Recovery =	198.72%#		
56) 1,1,2,2-Tetrachloroeth...	10.233	84	519355	82.28	ug/L	0.01
Spiked Amount 50.000	Range 65 - 120		Recovery =	164.56%#		
66) 1,2-Dichlorobenzene-d4	11.635	152	427846	87.76	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	175.52%#		
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	4208	1.07	ug/L	81
46) Tetrachloroethene	7.979	164	197296	54.32	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042621\
Data File : VV021150.D
Acq On : 26 Apr 2021 18:23
Operator : SY/MD
Sample : M2165-16ME
Misc : 5.56g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB3F7ME

Quant Time: Apr 27 04:18:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 27 04:16:04 2021
Response via : Initial Calibration

