

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042623\
 Data File : VW030875.D
 Acq On : 26 Apr 2023 13:08
 Operator : SY/MD
 Sample : VW0426MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK336

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/27/2023
 Supervised By :Mahesh Dadoda 04/27/2023

Quant Time: Apr 27 04:45:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 27 04:45:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	215061	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	209996	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	109159	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	63774	40.451	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	80.900%		
7) Chloroethane-d5	1.536	69	56759m	42.962	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.920%		
11) 1,1-Dichloroethene-d2	2.066	65	33258	44.483	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.960%		
21) 2-Butanone-d5	3.799	46	88632	102.353	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.350%		
24) Chloroform-d	4.259	84	134420	47.048	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.100%		
26) 1,2-Dichloroethane-d4	4.950	65	98453	53.598	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	107.200%		
32) Benzene-d6	4.970	84	249274	50.207	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.420%		
36) 1,2-Dichloropropane-d6	5.998	67	73016	47.345	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	94.680%		
41) Toluene-d8	7.252	98	226765	48.646	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.300%		
43) trans-1,3-Dichloroprop...	7.561	79	37557	48.317	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	96.640%		
47) 2-Hexanone-d5	8.034	63	54732	90.087	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	90.090%		
56) 1,1,2,2-Tetrachloroeth...	10.159	84	105148	46.986	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	93.980%		
66) 1,2-Dichlorobenzene-d4	11.567	152	95349	54.955	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	109.900%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042623\
Data File : VV030875.D
Acq On : 26 Apr 2023 13:08
Operator : SY/MD
Sample : VV0426MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK336

Quant Time: Apr 27 04:45:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 27 04:45:18 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/27/2023
Supervised By :Mahesh Dadoda 04/27/2023

