

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042624\
 Data File : VW035422.D
 Acq On : 26 Apr 2024 13:09
 Operator : SY/MD
 Sample : VW0426WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK277

Quant Time: Apr 27 00:33:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 27 00:18:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	522574	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	515226	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	225283	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	190814	45.435	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.880%
7) Chloroethane-d5	1.532	69	151643	63.022	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	126.040%
11) 1,1-Dichloroethene-d2	2.056	65	77891	45.379	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.760%
21) 2-Butanone-d5	3.793	46	194452	90.342	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.340%
24) Chloroform-d	4.246	84	315139	44.594	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.180%
26) 1,2-Dichloroethane-d4	4.937	65	195706	47.546	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.100%
32) Benzene-d6	4.957	84	646094	47.718	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.440%
36) 1,2-Dichloropropane-d6	5.985	67	203933	48.817	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.640%
41) Toluene-d8	7.239	98	553805	45.380	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.760%
43) trans-1,3-Dichloroprop...	7.551	79	84750	43.368	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.740%
47) 2-Hexanone-d5	8.024	63	102470	85.087	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.090%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	271959	45.676	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.360%
66) 1,2-Dichlorobenzene-d4	11.558	152	211347	53.621	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042624\
Data File : VV035422.D
Acq On : 26 Apr 2024 13:09
Operator : SY/MD
Sample : VV0426WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK277

Quant Time: Apr 27 00:33:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 27 00:18:54 2024
Response via : Initial Calibration

