

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042624\
 Data File : VW035454.D
 Acq On : 27 Apr 2024 02:47
 Operator : SY/MD
 Sample : P2279-14 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CORR7

Quant Time: Apr 27 05:29:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 27 01:20:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	436850	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	417701	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	170470	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	151660	43.199	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.400%	
7) Chloroethane-d5	1.529	69	101422	50.421	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	100.840%	
11) 1,1-Dichloroethene-d2	2.053	65	60955	42.481	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	84.960%	
21) 2-Butanone-d5	3.799	46	208247	115.737	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	115.740%	
24) Chloroform-d	4.246	84	283923	48.061	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	96.120%	
26) 1,2-Dichloroethane-d4	4.937	65	183507	53.331	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	106.660%	
32) Benzene-d6	4.956	84	549147	50.028	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.060%	
36) 1,2-Dichloropropane-d6	5.985	67	181285	53.528	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	107.060%	
41) Toluene-d8	7.243	98	433231	43.788	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.580%	
43) trans-1,3-Dichloroprop...	7.554	79	76128	48.051	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.100%	
47) 2-Hexanone-d5	8.030	63	101626	104.089	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	104.090%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	262853	54.454	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	108.900%	
66) 1,2-Dichlorobenzene-d4	11.561	152	158981	53.305	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042624\
Data File : VV035454.D
Acq On : 27 Apr 2024 02:47
Operator : SY/MD
Sample : P2279-14 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CORR7

Quant Time: Apr 27 05:29:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 27 01:20:16 2024
Response via : Initial Calibration

