

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042624\
 Data File : VW035427.D
 Acq On : 26 Apr 2024 15:16
 Operator : SY/MD
 Sample : P2269-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0RQ0

Quant Time: Apr 27 00:35:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 27 00:18:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	514189	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	497321	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	212547	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	184079	44.547	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.100%	
7) Chloroethane-d5	1.526	69	123336	52.093	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	104.180%	
11) 1,1-Dichloroethene-d2	2.056	65	76790	45.467	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.940%	
21) 2-Butanone-d5	3.793	46	201653	95.215	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	95.220%	
24) Chloroform-d	4.246	84	311871	44.851	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.700%	
26) 1,2-Dichloroethane-d4	4.940	65	192331	47.488	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.980%	
32) Benzene-d6	4.956	84	637068	48.746	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.500%	
36) 1,2-Dichloropropane-d6	5.985	67	198768	49.294	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.580%	
41) Toluene-d8	7.239	98	542347	46.041	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	92.080%	
43) trans-1,3-Dichloroprop...	7.551	79	85597	45.378	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.760%	
47) 2-Hexanone-d5	8.027	63	105462	90.724	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	90.720%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	266893	46.439	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	92.880%	
66) 1,2-Dichlorobenzene-d4	11.557	152	197594	53.136	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.280%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042624\
Data File : VV035427.D
Acq On : 26 Apr 2024 15:16
Operator : SY/MD
Sample : P2269-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0RQ0

Quant Time: Apr 27 00:35:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 27 00:18:54 2024
Response via : Initial Calibration

