

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042719\
 Data File : VV010537.D
 Acq On : 27 Apr 2019 08:34
 Operator : SY/MD
 Sample : K2586-15
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0X57

Quant Time: Apr 29 02:45:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 29 01:58:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	238187	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	242947	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	110225	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	52905	4.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.40%
7) Chloroethane-d5	1.57	69	39412	6.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	127.20%
11) 1,1-Dichloroethene-d2	2.13	63	74381	3.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.20%
20) 2-Butanone-d5	3.94	46	128030	42.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.70%
24) Chloroform-d	4.40	84	136231	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.60%
26) 1,2-Dichloroethane-d4	5.08	65	70145	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.10	84	291455	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.12	67	80756	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.36	98	268408	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
43) trans-1,3-Dichloropropene-	7.67	79	26610	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.13	63	128330	46.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.06%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	59446	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	119701	5.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.60%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.76	101	1290355	49.650 ug/L	97
13) Acetone	2.19	43	26757	17.975 ug/L	93
22) cis-1,2-Dichloroethene	3.96	96	4191	0.236 ug/L	85
31) Carbon tetrachloride	4.87	117	40289	1.481 ug/L	97
34) Trichloroethene	5.96	95	43241	2.364 ug/L	97
45) 1,1,2-Trichloroethane	7.89	97	2432	0.198 ug/L	90
47) Tetrachloroethene	8.02	164	5919723	346.375 ug/L	99
51) Chlorobenzene	8.93	112	15840	0.327 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042719\
Data File : VV010537.D
Acq On : 27 Apr 2019 08:34
Operator : SY/MD
Sample : K2586-15
Misc : 25mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0X57

Quant Time: Apr 29 02:45:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 29 01:58:06 2019
Response via : Initial Calibration

