

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042719\
 Data File : VV010516.D
 Acq On : 26 Apr 2019 21:14
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV39

Quant Time: Apr 26 22:42:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 22:39:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	115	0.00
2 T	Dichlorodifluoromethane	0.401	0.383	4.5	101	0.00
3 T	Chloromethane	0.266	0.237	10.9	104	0.00
4 S	Vinyl Chloride-d3	0.243	0.228	6.2	98	0.00
5 T	Vinyl chloride	0.278	0.266	4.3	103	0.00
6 T	Bromomethane	0.121	0.142	-17.4	112	0.00
7 S	Chloroethane-d5	0.130	0.116	10.8	102	0.00
8 T	Chloroethane	0.112	0.099	11.6	103	0.00
9 T	Trichlorofluoromethane	0.546	0.542	0.7	106	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.271	0.272	-0.4	106	0.00
11 S	1,1-Dichloroethene-d2	0.458	0.456	0.4	104	0.00
12 T	1,1-Dichloroethene	0.241	0.235	2.5	103	0.00
13 T	Acetone	0.031	0.028	9.7	91	0.01
14 T	Carbon disulfide	0.609	0.605	0.7	106	0.00
15 T	Methyl Acetate	0.104	0.091	12.5	101	0.00
16 T	Methylene chloride	0.365	0.335	8.2	102	0.00
17 T	Methyl tert-butyl Ether	0.699	0.671	4.0	101	0.00
18 T	trans-1,2-Dichloroethene	0.345	0.345	0.0	105	0.00
19 T	1,1-Dichloroethane	0.569	0.557	2.1	103	0.00
20 S	2-Butanone-d5	0.063	0.060	4.8	93	0.00
21 T	2-Butanone	0.065	0.060	7.7	95	0.00
22 T	cis-1,2-Dichloroethene	0.372	0.363	2.4	107	0.00
23 T	Bromochloromethane	0.174	0.167	4.0	102	0.00
24 S	Chloroform-d	0.652	0.645	1.1	102	0.00
25 T	Chloroform	0.645	0.628	2.6	103	0.00
26 S	1,2-Dichloroethane-d4	0.301	0.294	2.3	98	0.00
27 T	1,2-Dichloroethane	0.367	0.352	4.1	100	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
29 T	1,1,1-Trichloroethane	0.586	0.598	-2.0	106	0.00
30 T	Cyclohexane	0.424	0.459	-8.3	109	0.00
31 T	Carbon tetrachloride	0.560	0.573	-2.3	105	0.00
32 S	Benzene-d6	1.234	1.259	-2.0	99	0.00
33 T	Benzene	1.294	1.340	-3.6	105	0.00
34 T	Trichloroethene	0.377	0.384	-1.9	106	0.00
35 T	Methylcyclohexane	0.489	0.504	-3.1	109	0.00
36 S	1,2-Dichloropropane-d6	0.352	0.352	0.0	99	0.00
37 T	1,2-Dichloropropane	0.298	0.296	0.7	98	0.00
38 T	Bromodichloromethane	0.428	0.424	0.9	104	0.00
39 T	cis-1,3-Dichloropropene	0.431	0.450	-4.4	105	0.00
40 T	4-Methyl-2-pentanone	0.158	0.155	1.9	92	0.00
41 S	Toluene-d8	1.180	1.257	-6.5	100	0.00
42 T	Toluene	1.420	1.516	-6.8	104	0.00
43 S	trans-1,3-Dichloropropene-d	0.134	0.139	-3.7	105	0.00
44 T	trans-1,3-Dichloropropene	0.339	0.353	-4.1	104	0.00
45 T	1,1,2-Trichloroethane	0.253	0.249	1.6	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042719\
 Data File : VV010516.D
 Acq On : 26 Apr 2019 21:14
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV39

Quant Time: Apr 26 22:42:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 26 22:39:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 S	2-Hexanone-d5	0.057	0.057	0.0	92	0.00
47 T	Tetrachloroethene	0.352	0.357	-1.4	105	0.00
48 T	2-Hexanone	0.110	0.112	-1.8	93	0.00
49 T	Dibromochloromethane	0.311	0.312	-0.3	99	0.00
50 T	1,2-Dibromoethane	0.234	0.241	-3.0	102	0.00
51 T	Chlorobenzene	0.997	1.033	-3.6	106	0.00
52 T	Ethylbenzene	1.496	1.591	-6.4	105	0.00
53 T	m,p-xylene	0.588	0.653	-11.1	110	0.00
54 T	o-xylene	0.575	0.630	-9.6	107	0.00
55 T	Styrene	0.950	1.060	-11.6	105	0.00
56 T	Isopropylbenzene	1.506	1.668	-10.8	108	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.279	0.266	4.7	92	0.00
58 T	1,1,2,2-Tetrachloroethane	0.276	0.272	1.4	96	0.00
59	1,2,3-Trichloropropane	0.191	0.191	0.0	98	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
61 T	Bromoform	0.325	0.330	-1.5	103	0.00
62 T	1,3-Dichlorobenzene	1.542	1.631	-5.8	107	0.00
63 T	1,4-Dichlorobenzene	1.582	1.664	-5.2	108	0.00
64 S	1,2-Dichlorobenzene-d4	0.964	1.018	-5.6	104	0.00
65 T	1,2-Dichlorobenzene	1.519	1.581	-4.1	105	0.00
66 T	1,2-Dibromo-3-chloropropane	0.084	0.076	9.5	90	0.00
67	1,3,5-Trichlorobenzene	1.293	1.405	-8.7	110	0.00
68 T	1,2,4-trichlorobenzene	0.987	1.087	-10.1	111	0.00
69	Naphthalene	1.468	1.556	-6.0	106	0.00
70 T	1,2,3-Trichlorobenzene	0.982	1.045	-6.4	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0