

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042720\
 Data File : VV015721.D
 Acq On : 27 Apr 2020 12:04
 Operator : SY/MD
 Sample : L2416-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKP0

Quant Time: Apr 28 02:58:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 28 02:16:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	135785	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	132369	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	63106	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	25323	3.57	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.40%
7) Chloroethane-d5	1.58	69	28537	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.40%
11) 1,1-Dichloroethene-d2	2.12	63	43844	2.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.40%#
20) 2-Butanone-d5	3.91	46	109262	47.95	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.90%
24) Chloroform-d	4.38	84	74417	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.00%
26) 1,2-Dichloroethane-d4	5.06	65	43529	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	5.08	84	137266	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.80%
36) 1,2-Dichloropropane-d6	6.09	67	45240	4.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.60%
41) Toluene-d8	7.34	98	122669	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.80%
43) trans-1,3-Dichloropropene-	7.65	79	16605	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.11	63	77351	40.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.26%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	34059	4.53	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	48645	4.53	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	4184	2.645	ug/L 98
22) cis-1,2-Dichloroethene	3.94	96	1599	0.163	ug/L 93
33) Benzene	5.13	78	15533	0.411	ug/L 100
42) Toluene	7.41	91	21507	0.532	ug/L 95
53) m,p-xylene	9.17	106	1994	0.116	ug/L 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042720\
Data File : VV015721.D
Acq On : 27 Apr 2020 12:04
Operator : SY/MD
Sample : L2416-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKPO

Quant Time: Apr 28 02:58:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 28 02:16:08 2020
Response via : Initial Calibration

