

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042720\
 Data File : VV015736.D
 Acq On : 27 Apr 2020 17:57
 Operator : SY/MD
 Sample : L2413-19DL 100X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETKM8DL

Quant Time: Apr 28 06:39:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 28 02:16:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	126093	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	122040	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	57230	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	25914	3.93	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.60%
7) Chloroethane-d5	1.58	69	27783	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.12	63	45479	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	3.93	46	116119	54.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.76%
24) Chloroform-d	4.38	84	71941	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	5.06	65	45407	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.08	84	141259	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.09	67	46133	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.34	98	126028	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
43) trans-1,3-Dichloropropene-	7.65	79	13838	3.84	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.80%
46) 2-Hexanone-d5	8.11	63	84385	48.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.16%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	34153	4.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	51509	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	19023	1.906	ug/L # 79
16) Methylene chloride	2.52	84	4087	0.451	ug/L 87
18) trans-1,2-Dichloroethene	2.78	96	1022	0.124	ug/L 84
22) cis-1,2-Dichloroethene	3.94	96	113391	12.480	ug/L 96
34) Trichloroethene	5.94	95	7077	0.735	ug/L 98
47) Tetrachloroethene	8.00	164	12989	1.873	ug/L 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042720\
Data File : VV015736.D
Acq On : 27 Apr 2020 17:57
Operator : SY/MD
Sample : L2413-19DL 100X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETKM8DL

Quant Time: Apr 28 06:39:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 28 02:16:08 2020
Response via : Initial Calibration

