

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042721\
 Data File : VV021164.D
 Acq On : 27 Apr 2021 10:35
 Operator : SY/MD
 Sample : VV0427MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK173

Quant Time: Apr 28 03:28:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 28 03:28:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	669085	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	644894	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	306861	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	260036	47.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.60%
7) Chloroethane-d5	1.571	69	196496	48.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.16%
11) 1,1-Dichloroethene-d2	2.108	63	361901	35.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.38%
21) 2-Butanone-d5	3.899	46	263974	98.32	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.32%
24) Chloroform-d	4.352	84	467503	47.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.034	65	314017	50.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.80%
32) Benzene-d6	5.050	84	855522	49.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.62%
36) 1,2-Dichloropropane-d6	6.072	67	251636	49.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.16%
41) Toluene-d8	7.317	98	799224	48.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.16%
43) trans-1,3-Dichloroprop...	7.625	79	136102	48.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.16%
47) 2-Hexanone-d5	8.091	63	195027	95.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.92%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	355716	46.00	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.00%
66) 1,2-Dichlorobenzene-d4	11.632	152	328857	54.25	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.50%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042721\
Data File : VV021164.D
Acq On : 27 Apr 2021 10:35
Operator : SY/MD
Sample : VV0427MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK173

Quant Time: Apr 28 03:28:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 28 03:28:03 2021
Response via : Initial Calibration

