

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042723\
 Data File : VW030903.D
 Acq On : 27 Apr 2023 16:02
 Operator : SY/MD
 Sample : 02447-16ME
 Misc : 6.08g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Apr 28 00:52:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 27 04:45:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	220801	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	215566	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	110636	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	68996	42.626	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.260%
7) Chloroethane-d5	1.535	69	56169	41.410	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.820%
11) 1,1-Dichloroethene-d2	2.066	65	33335	43.427	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.860%
21) 2-Butanone-d5	3.796	46	88020	99.004	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.000%
24) Chloroform-d	4.259	84	132462	45.157	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.320%
26) 1,2-Dichloroethane-d4	4.950	65	98309	52.128	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.260%
32) Benzene-d6	4.969	84	244967	48.065	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.120%
36) 1,2-Dichloropropane-d6	5.998	67	72171	45.587	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.180%
41) Toluene-d8	7.252	98	218902	45.746	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.500%
43) trans-1,3-Dichloroprop...	7.561	79	38544	48.305	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.620%
47) 2-Hexanone-d5	8.033	63	57617	92.386	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.390%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	105127	45.763	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.520%
66) 1,2-Dichlorobenzene-d4	11.567	152	93070	52.925	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042723\
Data File : VV030903.D
Acq On : 27 Apr 2023 16:02
Operator : SY/MD
Sample : 02447-16ME
Misc : 6.08g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 28 00:52:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 27 04:45:18 2023
Response via : Initial Calibration

