

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042723\
 Data File : VW030904.D
 Acq On : 27 Apr 2023 16:25
 Operator : SY/MD
 Sample : 02447-21ME
 Misc : 4.55g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW939ME

Quant Time: Apr 28 00:52:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 27 04:45:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	213499	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	208595	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	107757	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	66797	42.679	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.360%
7) Chloroethane-d5	1.536	69	53624	40.886	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.780%
11) 1,1-Dichloroethene-d2	2.063	65	32108	43.259	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.520%
21) 2-Butanone-d5	3.796	46	87348	101.608	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.610%
24) Chloroform-d	4.262	84	128210	45.202	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.400%
26) 1,2-Dichloroethane-d4	4.950	65	95574	52.411	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.820%
32) Benzene-d6	4.970	84	234116	47.471	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.940%
36) 1,2-Dichloropropane-d6	5.998	67	71173	46.459	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.920%
41) Toluene-d8	7.252	98	212824	45.962	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.920%
43) trans-1,3-Dichloroprop...	7.561	79	35378	45.819	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.640%
47) 2-Hexanone-d5	8.034	63	55943	92.699	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.700%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	101648	45.727	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.460%
66) 1,2-Dichlorobenzene-d4	11.571	152	92188	53.824	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042723\
Data File : VV030904.D
Acq On : 27 Apr 2023 16:25
Operator : SY/MD
Sample : 02447-21ME
Misc : 4.55g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW939ME

Quant Time: Apr 28 00:52:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041723WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 27 04:45:18 2023
Response via : Initial Calibration

