

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042820\
 Data File : VV015768.D
 Acq On : 29 Apr 2020 00:16
 Operator : SY/MD
 Sample : L2419-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF4

Quant Time: Apr 29 05:26:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 29 03:36:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	120987	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	117980	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	55736	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	37955	6.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.00%
7) Chloroethane-d5	1.58	69	33056	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	2.12	63	54661	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.00%
20) 2-Butanone-d5	3.92	46	86305	42.51	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.02%
24) Chloroform-d	4.38	84	66036	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
26) 1,2-Dichloroethane-d4	5.06	65	38469	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.40%
32) Benzene-d6	5.08	84	134693	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
36) 1,2-Dichloropropane-d6	6.10	67	41057	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.34	98	118890	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
43) trans-1,3-Dichloropropene-	7.65	79	13241	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	8.11	63	60013	35.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.74%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	26752	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	42738	4.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.20%

Target Compounds					Ovalue
3) Chloromethane	1.25	50	1504	0.168 ug/L	91
13) Acetone	2.19	43	7785	5.522 ug/L	100
25) Chloroform	4.41	83	4178	0.254 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042820\
Data File : VV015768.D
Acq On : 29 Apr 2020 00:16
Operator : SY/MD
Sample : L2419-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF4

Quant Time: Apr 29 05:26:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 29 03:36:29 2020
Response via : Initial Calibration

