

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042820\
 Data File : VV015773.D
 Acq On : 29 Apr 2020 02:14
 Operator : SY/MD
 Sample : L2419-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF1

Quant Time: Apr 29 06:02:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 29 03:36:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	114605	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	109478	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	48304	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	37323	6.23	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	124.60%
7) Chloroethane-d5	1.58	69	32811	5.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.60%
11) 1,1-Dichloroethene-d2	2.12	63	53098	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	3.91	46	84439	43.91	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.82%
24) Chloroform-d	4.38	84	57180	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
26) 1,2-Dichloroethane-d4	5.06	65	38471	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.08	84	130043	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
36) 1,2-Dichloropropane-d6	6.10	67	40000	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.80%
41) Toluene-d8	7.34	98	112111	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	7.65	79	12477	3.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.20%
46) 2-Hexanone-d5	8.11	63	64275	40.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.64%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	26525	4.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	36365	4.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	4254	3.186	ug/L 97
14) Carbon disulfide	2.31	76	2440	0.129	ug/L # 87
22) cis-1,2-Dichloroethene	3.94	96	62726	7.596	ug/L 96
25) Chloroform	4.40	83	11693	0.751	ug/L 95
29) 1,1,1-Trichloroethane	4.63	97	4844	0.349	ug/L 100
34) Trichloroethene	5.94	95	19534	2.261	ug/L 99
47) Tetrachloroethene	8.02	164	15628738	2512.830	ug/L 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042820\
Data File : VV015773.D
Acq On : 29 Apr 2020 02:14
Operator : SY/MD
Sample : L2419-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF1

Quant Time: Apr 29 06:02:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 29 03:36:29 2020
Response via : Initial Calibration

