

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
 Data File : VV021191.D
 Acq On : 28 Apr 2021 11:11
 Operator : SY/MD
 Sample : VV0428MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK175

Quant Time: Apr 29 04:13:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 29 04:12:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	837337	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	815653	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	391864	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	293074	42.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.20%
7) Chloroethane-d5	1.571	69	235535	46.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.06%
11) 1,1-Dichloroethene-d2	2.108	63	397453	31.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.64%
21) 2-Butanone-d5	3.905	46	302074	89.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.91%
24) Chloroform-d	4.349	84	552655	45.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.10%
26) 1,2-Dichloroethane-d4	5.034	65	344797	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.44%
32) Benzene-d6	5.050	84	1055377	48.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.16%
36) 1,2-Dichloropropane-d6	6.072	67	329099	50.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.52%
41) Toluene-d8	7.317	98	995697	47.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.72%
43) trans-1,3-Dichloroprop...	7.625	79	162179	45.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.54%
47) 2-Hexanone-d5	8.092	63	220301	85.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.66%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	445447	45.55	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.10%
66) 1,2-Dichlorobenzene-d4	11.632	152	406424	52.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
Data File : VV021191.D
Acq On : 28 Apr 2021 11:11
Operator : SY/MD
Sample : VV0428MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK175

Quant Time: Apr 29 04:13:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 29 04:12:32 2021
Response via : Initial Calibration

