

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
 Data File : VV021196.D
 Acq On : 28 Apr 2021 13:07
 Operator : SY/MD
 Sample : VIBLK173
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK173

Quant Time: Apr 29 04:13:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 29 04:12:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	683801	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	695198	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	330429	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	223635	39.80	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	79.60%		
7) Chloroethane-d5	1.571	69	189276	45.79	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	91.58%		
11) 1,1-Dichloroethene-d2	2.111	63	317037	30.59	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	61.18%		
21) 2-Butanone-d5	3.905	46	300714	109.60	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	109.60%		
24) Chloroform-d	4.355	84	487937	48.71	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	97.42%		
26) 1,2-Dichloroethane-d4	5.037	65	320505	50.33	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.66%		
32) Benzene-d6	5.056	84	928648	50.16	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	100.32%		
36) 1,2-Dichloropropane-d6	6.076	67	292975	53.01	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	106.02%		
41) Toluene-d8	7.320	98	848754	47.37	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	94.74%		
43) trans-1,3-Dichloroprop...	7.625	79	144380	47.81	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.62%		
47) 2-Hexanone-d5	8.095	63	230190	105.02	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	105.02%		
56) 1,1,2,2-Tetrachloroeth...	10.223	84	427518	51.29	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	102.58%		
66) 1,2-Dichlorobenzene-d4	11.632	152	362517	55.54	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	111.08%		
Target Compounds						Qvalue
46) Tetrachloroethene	7.982	164	23339	4.87	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
Data File : VV021196.D
Acq On : 28 Apr 2021 13:07
Operator : SY/MD
Sample : VIBLK173
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK173

Quant Time: Apr 29 04:13:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 29 04:12:32 2021
Response via : Initial Calibration

