

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042821\
 Data File : VW021209.D
 Acq On : 28 Apr 2021 18:35
 Operator : SY/MD
 Sample : M2165-21ME 5X
 Misc : 6.48g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB3G2ME

Quant Time: Apr 29 04:15:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 29 04:12:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	792605	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	793047	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	402827	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	224574	34.48	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	68.96%		
7) Chloroethane-d5	1.568	69	181442	37.87	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	75.74%		
11) 1,1-Dichloroethene-d2	2.105	63	320969	26.72	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	53.44%#		
21) 2-Butanone-d5	3.902	46	288925	90.85	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	90.85%		
24) Chloroform-d	4.352	84	474367	40.86	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	81.72%		
26) 1,2-Dichloroethane-d4	5.034	65	300754	40.75	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	81.50%		
32) Benzene-d6	5.053	84	918056	43.47	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.94%		
36) 1,2-Dichloropropane-d6	6.072	67	286446	45.44	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	90.88%		
41) Toluene-d8	7.320	98	861044	42.12	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	84.24%		
43) trans-1,3-Dichloroprop...	7.625	79	138963	40.33	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	80.66%		
47) 2-Hexanone-d5	8.095	63	225299	90.10	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	90.10%		
56) 1,1,2,2-Tetrachloroeth...	10.223	84	404988	42.59	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.18%		
66) 1,2-Dichlorobenzene-d4	11.632	152	386252	48.54	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.08%		
Target Compounds						Qvalue
16) Methylene chloride	2.507	84	8012	1.36	ug/L	97
46) Tetrachloroethene	7.979	164	569958	104.15	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
Data File : VV021209.D
Acq On : 28 Apr 2021 18:35
Operator : SY/MD
Sample : M2165-21ME 5X
Misc : 6.48g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB3G2ME

Quant Time: Apr 29 04:15:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 29 04:12:32 2021
Response via : Initial Calibration

