

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
 Data File : VV021212.D
 Acq On : 28 Apr 2021 19:46
 Operator : SY/MD
 Sample : M2166-03ME
 Misc : 5.48g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GB3G6ME

Quant Time: Apr 29 04:15:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 29 04:12:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	917115	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.860	117	916729	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.259	152	459064	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	189438	25.14	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	50.28%#		
7) Chloroethane-d5	1.568	69	138505	24.98	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	49.96%#		
11) 1,1-Dichloroethene-d2	2.085	63	256188	18.43	ug/L	-0.03
Spiked Amount 50.000	Range 60 - 125		Recovery =	36.86%#		
21) 2-Butanone-d5	3.915	46	193541	52.59	ug/L	0.02
Spiked Amount 100.000	Range 40 - 130		Recovery =	52.59%		
24) Chloroform-d	4.349	84	427345	31.81	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	63.62%#		
26) 1,2-Dichloroethane-d4	5.034	65	280417	32.83	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	65.66%#		
32) Benzene-d6	5.047	84	834753	34.19	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	68.38%#		
36) 1,2-Dichloropropane-d6	6.072	67	256572	35.21	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	70.42%		
41) Toluene-d8	7.320	98	777087	32.89	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	65.78%#		
43) trans-1,3-Dichloroprop...	7.628	79	128614	32.29	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	64.58%		
47) 2-Hexanone-d5	8.127	63	193082	66.80	ug/L	0.04
Spiked Amount 100.000	Range 45 - 130		Recovery =	66.80%		
56) 1,1,2,2-Tetrachloroeth...	10.230	84	355516	32.34	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	64.68%#		
66) 1,2-Dichlorobenzene-d4	11.635	152	339672	37.46	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	74.92%#		
Target Compounds						Qvalue
16) Methylene chloride	2.497	84	11940	1.75	ug/L	80
46) Tetrachloroethene	7.979	164	222464	35.17	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042821\
Data File : VV021212.D
Acq On : 28 Apr 2021 19:46
Operator : SY/MD
Sample : M2166-03ME
Misc : 5.48g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GB3G6ME

Quant Time: Apr 29 04:15:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042121WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 29 04:12:32 2021
Response via : Initial Calibration

