

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042823\
 Data File : VW030929.D
 Acq On : 28 Apr 2023 18:32
 Operator : SY/MD
 Sample : 02488-18DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHE73DL

Quant Time: Apr 29 01:11:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 29 01:06:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	128699	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	125916	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	62151	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	46299	4.366	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	87.400%	
7) Chloroethane-d5	1.536	69	41377	4.784	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.063	65	24851	4.428	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	88.600%	
20) 2-Butanone-d5	3.838	46	80861	43.004	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.000%	
24) Chloroform-d	4.259	84	94934	4.505	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
26) 1,2-Dichloroethane-d4	4.953	65	47017	4.389	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.800%	
32) Benzene-d6	4.969	84	166614	4.514	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.200%	
36) 1,2-Dichloropropane-d6	5.998	67	48043	4.477	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.600%	
41) Toluene-d8	7.252	98	141506	4.026	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.600%	
43) trans-1,3-Dichloroprop...	7.564	79	17206	4.277	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.600%	
46) 2-Hexanone-d5	8.037	63	59179	41.986	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	83.980%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	33259	4.275	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.600%	
66) 1,2-Dichlorobenzene-d4	11.570	152	54870	4.736	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.800%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.072	96	12571	1.451	ug/L #	1
14) Carbon disulfide	2.246	76	23042	1.017	ug/L	98
16) Methylene chloride	2.455	84	4090	0.417	ug/L	84
29) 1,1,1-Trichloroethane	4.526	97	5057	0.320	ug/L #	82
34) Trichloroethene	5.844	95	39768	4.710	ug/L	97
42) Toluene	7.326	91	29584	0.868	ug/L	95
47) Tetrachloroethene	7.911	164	11697	1.452	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042823\
Data File : VV030929.D
Acq On : 28 Apr 2023 18:32
Operator : SY/MD
Sample : 02488-18DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHE73DL

Quant Time: Apr 29 01:11:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 29 01:06:38 2023
Response via : Initial Calibration

