

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042823\
 Data File : VW030923.D
 Acq On : 28 Apr 2023 16:09
 Operator : SY/MD
 Sample : 02488-16
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHE71

Quant Time: Apr 29 01:09:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 29 01:06:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	121907	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	121689	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	59981	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	50185	4.997	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.000%
7) Chloroethane-d5	1.535	69	43288	5.283	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.600%
11) 1,1-Dichloroethene-d2	2.063	65	26466	4.979	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.600%
20) 2-Butanone-d5	3.838	46	71022	39.876	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.760%
24) Chloroform-d	4.259	84	99700	4.995	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	4.953	65	52865	5.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.200%
32) Benzene-d6	4.969	84	176428	4.946	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	5.998	67	50793	4.897	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.252	98	157455	4.635	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
43) trans-1,3-Dichloroprop...	7.564	79	19467	5.007	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.200%
46) 2-Hexanone-d5	8.037	63	61599	45.221	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.440%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	36402	4.842	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.800%
66) 1,2-Dichlorobenzene-d4	11.570	152	60200	5.384	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042823\
Data File : VV030923.D
Acq On : 28 Apr 2023 16:09
Operator : SY/MD
Sample : 02488-16
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHE71

Quant Time: Apr 29 01:09:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 29 01:06:38 2023
Response via : Initial Calibration

