

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042919\
 Data File : VV010542.D
 Acq On : 29 Apr 2019 11:15
 Operator : SY/MD
 Sample : VSTDCCC0.5
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.510

Quant Time: Apr 30 01:23:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 27 04:44:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	7449	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	6775	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.30	152	3072	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2467	0.45	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	90.00%
4) 1,2-Dichloroethane-d4	5.07	65	2429	0.48	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.00%
7) Toluene-d8	7.35	98	8008	0.57	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	114.00%
9) 1,1,2,2-Tetrachloroethane-	10.25	84	1850	0.47	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	94.00%
11) 1-2-dichlorobenzene-d4	11.67	152	3150	0.51	ug/L	0.00
Spiked Amount	0.500	Range	80 - 120	Recovery	=	102.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	2338	0.420 ug/L	99
6) Trichloroethene	5.96	95	2847	0.498 ug/L	99
8) 1,2-Dibromoethane	8.40	107	1615	0.492 ug/L	100
12) 1,2,3-Trichloropropane	10.30	75	1367	0.521 ug/L	97
13) 1,2-Dibromo-3-chloropropan	12.46	75	281	0.393 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042919\
Data File : VV010542.D
Acq On : 29 Apr 2019 11:15
Operator : SY/MD
Sample : VSTDCCC0.5
Misc : 25mL/MSVOA V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.510

Quant Time: Apr 30 01:23:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042619SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 27 04:44:38 2019
Response via : Initial Calibration

