

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV042920\
 Data File : VV015786.D
 Acq On : 29 Apr 2020 12:44
 Operator : SY/MD
 Sample : L2419-05DL 400X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF2DL

Quant Time: Apr 30 05:08:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 30 01:35:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	126170	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	122709	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	57474	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	38962	5.91	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	118.20%
7) Chloroethane-d5	1.58	69	33617	5.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.80%
11) 1,1-Dichloroethene-d2	2.12	63	57559	3.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.80%
20) 2-Butanone-d5	3.96	46	94820	44.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.58%
24) Chloroform-d	4.38	84	66847	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
26) 1,2-Dichloroethane-d4	5.06	65	42232	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.07	84	139399	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.10	67	43945	4.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.00%
41) Toluene-d8	7.34	98	123876	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	7.65	79	15110	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.12	63	65775	37.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.54%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	29051	4.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	44379	4.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.80%

Target Compounds					Ovalue
47) Tetrachloroethene	8.00	164	37713	5.410 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV042920\
Data File : VV015786.D
Acq On : 29 Apr 2020 12:44
Operator : SY/MD
Sample : L2419-05DL 400X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF2DL

Quant Time: Apr 30 05:08:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 30 01:35:08 2020
Response via : Initial Calibration

