

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW042924\
 Data File : VW035467.D
 Acq On : 29 Apr 2024 13:45
 Operator : SY/MD
 Sample : P2279-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0RQ6

Quant Time: Apr 30 02:11:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 30 02:09:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	505864	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	504628	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	215005	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	172612	42.459	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.920%
7) Chloroethane-d5	1.526	69	115815	49.722	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.440%
11) 1,1-Dichloroethene-d2	2.053	65	74230	44.675	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.340%
21) 2-Butanone-d5	3.799	46	220644	105.897	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.900%
24) Chloroform-d	4.246	84	316515	46.268	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.540%
26) 1,2-Dichloroethane-d4	4.940	65	198713	49.872	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.740%
32) Benzene-d6	4.956	84	639026	48.188	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.380%
36) 1,2-Dichloropropane-d6	5.985	67	205670	50.267	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.540%
41) Toluene-d8	7.243	98	546844	45.751	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.500%
43) trans-1,3-Dichloroprop...	7.554	79	87202	45.559	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.120%
47) 2-Hexanone-d5	8.030	63	123672	104.849	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.850%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	289916	49.714	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.420%
66) 1,2-Dichlorobenzene-d4	11.557	152	204441	54.349	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV042924\
Data File : VV035467.D
Acq On : 29 Apr 2024 13:45
Operator : SY/MD
Sample : P2279-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0RQ6

Quant Time: Apr 30 02:11:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM042624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 30 02:09:05 2024
Response via : Initial Calibration

