

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043019\
 Data File : VV010552.D
 Acq On : 29 Apr 2019 17:40
 Operator : SY/MD
 Sample : K2586-17DL 5X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0X59DL

Quant Time: Apr 30 06:59:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 30 06:45:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	218624	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	217143	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96324	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	48681	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	1.57	69	32072	5.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.80%
11) 1,1-Dichloroethene-d2	2.13	63	67234	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	3.94	46	123937	45.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.38%
24) Chloroform-d	4.40	84	141753	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	5.08	65	65454	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	5.10	84	262985	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.12	67	75718	4.96	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.20%
41) Toluene-d8	7.36	98	249495	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
43) trans-1,3-Dichloropropene-	7.67	79	26945	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.13	63	111244	45.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.26%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	54811	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	101745	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.77	101	24155	1.013 ug/L	92
13) Acetone	2.19	43	5578	4.083 ug/L	98
34) Trichloroethene	5.96	95	4541	0.278 ug/L #	82
47) Tetrachloroethene	8.02	164	118268	7.742 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV043019\
Data File : VV010552.D
Acq On : 29 Apr 2019 17:40
Operator : SY/MD
Sample : K2586-17DL 5X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0X59DL

Quant Time: Apr 30 06:59:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 30 06:45:30 2019
Response via : Initial Calibration

