

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043019\
 Data File : VV010567.D
 Acq On : 30 Apr 2019 00:00
 Operator : SY/MD
 Sample : K2590-08
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0X99

Quant Time: Apr 30 07:46:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 30 06:45:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	216657	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	220516	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	98789	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	49347	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.57	69	32473	5.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.20%
11) 1,1-Dichloroethene-d2	2.13	63	67658	3.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.20%
20) 2-Butanone-d5	3.93	46	123161	45.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.62%
24) Chloroform-d	4.40	84	125346	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
26) 1,2-Dichloroethane-d4	5.08	65	67255	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.10	84	265990	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.12	67	77279	4.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.60%
41) Toluene-d8	7.36	98	251894	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	7.66	79	26991	4.55	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.00%
46) 2-Hexanone-d5	8.13	63	115831	46.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.54%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	57571	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	108719	5.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.20%

Target Compounds

					Ovalue
13) Acetone	2.19	43	17685	13.061 ug/L	97
34) Trichloroethene	5.96	95	62692	3.775 ug/L	98
42) Toluene	7.43	91	8754	0.140 ug/L	90
47) Tetrachloroethene	8.02	164	2970	0.191 ug/L	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV043019\
Data File : VV010567.D
Acq On : 30 Apr 2019 00:00
Operator : SY/MD
Sample : K2590-08
Misc : 25mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0X99

Quant Time: Apr 30 07:46:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 30 06:45:30 2019
Response via : Initial Calibration

