

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043019\
 Data File : VV010570.D
 Acq On : 30 Apr 2019 01:16
 Operator : SY/MD
 Sample : K2590-14
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XC0

Quant Time: Apr 30 07:54:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 30 06:45:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	211351	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	209290	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	96766	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	47384	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	1.57	69	32042	5.82	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	116.40%
11) 1,1-Dichloroethene-d2	2.13	63	64332	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	3.94	46	124833	47.08	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.16%
24) Chloroform-d	4.40	84	130255	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
26) 1,2-Dichloroethane-d4	5.08	65	67194	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
32) Benzene-d6	5.10	84	260700	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.12	67	76539	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.36	98	245170	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloropropene-	7.66	79	26299	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.13	63	113783	47.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.78%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	54841	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	107352	5.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.20%

Target Compounds

					Ovalue
9) Trichlorofluoromethane	1.77	101	5010	0.217 ug/L	95
13) Acetone	2.19	43	14509	10.985 ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	12755	0.810 ug/L	95
34) Trichloroethene	5.96	95	297070	18.850 ug/L	98
47) Tetrachloroethene	8.02	164	26843	1.823 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV043019\
Data File : VV010570.D
Acq On : 30 Apr 2019 01:16
Operator : SY/MD
Sample : K2590-14
Misc : 25mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XC0

Quant Time: Apr 30 07:54:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 30 06:45:30 2019
Response via : Initial Calibration

