

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043019\
 Data File : VV010554.D
 Acq On : 29 Apr 2019 18:31
 Operator : SY/MD
 Sample : K2586-22
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Apr 30 07:05:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 30 06:45:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	215189	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	219582	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	96476	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	50798	4.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.00%
7) Chloroethane-d5	1.58	69	25216	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.00%
11) 1,1-Dichloroethene-d2	2.13	63	70195	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	3.95	46	125379	46.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.88%
24) Chloroform-d	4.40	84	138777	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.08	65	68325	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.10	84	275979	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.12	67	78520	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.36	98	253927	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
43) trans-1,3-Dichloropropene-	7.67	79	27839	4.71	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.20%
46) 2-Hexanone-d5	8.14	63	108577	43.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	54587	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	99892	5.37	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV043019\
Data File : VV010554.D
Acq On : 29 Apr 2019 18:31
Operator : SY/MD
Sample : K2586-22
Misc : 25mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Apr 30 07:05:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 30 06:45:30 2019
Response via : Initial Calibration

