

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043020\
 Data File : VV015833.D
 Acq On : 30 Apr 2020 13:42
 Operator : SY/MD
 Sample : L2421-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFS4

Quant Time: May 02 05:08:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 01 01:42:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	122254	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	122002	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	57180	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	35061	5.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.80%
7) Chloroethane-d5	1.58	69	32200	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.12	63	52630	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.20%
20) 2-Butanone-d5	3.96	46	99028	48.27	ug/L	0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.54%
24) Chloroform-d	4.38	84	67559	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.06	65	42710	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.08	84	137788	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.10	67	43890	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.34	98	123582	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.65	79	14350	3.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.80%
46) 2-Hexanone-d5	8.12	63	69075	39.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.74%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	29226	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	44738	4.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.00%

Target Compounds					Ovalue
34) Trichloroethene	5.95	95	1640	0.170 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV043020\
Data File : VV015833.D
Acq On : 30 Apr 2020 13:42
Operator : SY/MD
Sample : L2421-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFS4

Quant Time: May 02 05:08:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 01 01:42:43 2020
Response via : Initial Calibration

