

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043020\
 Data File : VV015836.D
 Acq On : 30 Apr 2020 14:51
 Operator : SY/MD
 Sample : L2421-07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFSAS

Quant Time: May 01 03:03:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 01 01:42:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	118385	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	115190	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	51720	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	35612	5.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.20%
7) Chloroethane-d5	1.58	69	32693	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.60%
11) 1,1-Dichloroethene-d2	2.12	63	53712	3.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.40%
20) 2-Butanone-d5	3.94	46	100078	50.38	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.76%
24) Chloroform-d	4.38	84	71845	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.06	65	42206	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.08	84	139860	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.10	67	43086	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.34	98	122988	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.20%
43) trans-1,3-Dichloropropene-	7.65	79	14711	4.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.60%
46) 2-Hexanone-d5	8.12	63	68476	41.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.66%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	29696	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	44307	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	3.94	96	32704	3.834 ug/L	95
34) Trichloroethene	5.94	95	291961	32.124 ug/L	99
47) Tetrachloroethene	8.00	164	1665	0.254 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV043020\
Data File : VV015836.D
Acq On : 30 Apr 2020 14:51
Operator : SY/MD
Sample : L2421-07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFSAS

Quant Time: May 01 03:03:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 01 01:42:43 2020
Response via : Initial Calibration

