

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010588.D
 Acq On : 30 Apr 2019 21:40
 Operator : SY/MD
 Sample : K2590-10
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XA2

Quant Time: May 01 08:52:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 01 08:14:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	182576	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	186725	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	75513	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	55362	5.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.80%
7) Chloroethane-d5	1.56	69	38856	5.60	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.00%
11) 1,1-Dichloroethene-d2	2.12	63	77532	3.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.00%
20) 2-Butanone-d5	3.93	46	145313	62.78	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.56%
24) Chloroform-d	4.40	84	106958	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	5.08	65	61797	5.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.00%
32) Benzene-d6	5.10	84	227196	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
36) 1,2-Dichloropropane-d6	6.12	67	71510	5.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.40%
41) Toluene-d8	7.36	98	210188	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.20%
43) trans-1,3-Dichloropropene-	7.66	79	23138	4.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.60%
46) 2-Hexanone-d5	8.13	63	126929	65.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	130.30%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	52685	5.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	79751	6.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	121.60%#

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Trichlorofluoromethane	1.76	101	51332	2.140	ug/L	99
13) Acetone	2.19	43	42504	29.702	ug/L	99
34) Trichloroethene	5.96	95	6767	0.482	ug/L #	82
47) Tetrachloroethene	8.02	164	52420	4.897	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050119\
Data File : VV010588.D
Acq On : 30 Apr 2019 21:40
Operator : SY/MD
Sample : K2590-10
Misc : 25mL/MSVOA V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XA2

Quant Time: May 01 08:52:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 01 08:14:23 2019
Response via : Initial Calibration

