

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010590.D
 Acq On : 30 Apr 2019 22:31
 Operator : SY/MD
 Sample : K2590-18DL 25X
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XC5DL

Quant Time: May 01 08:57:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 01 08:14:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	176710	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	180346	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	73005	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	54418	5.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.60%
7) Chloroethane-d5	1.58	69	26459	3.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.80%
11) 1,1-Dichloroethene-d2	2.13	63	80621	4.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.00%
20) 2-Butanone-d5	3.95	46	150055	66.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	133.98%#
24) Chloroform-d	4.40	84	123417	5.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.80%
26) 1,2-Dichloroethane-d4	5.08	65	64157	6.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.20%
32) Benzene-d6	5.10	84	246233	6.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	121.40%
36) 1,2-Dichloropropane-d6	6.12	67	74228	6.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	121.80%
41) Toluene-d8	7.36	98	211464	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
43) trans-1,3-Dichloropropene-	7.67	79	25457	5.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.80%
46) 2-Hexanone-d5	8.14	63	126903	67.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	134.88%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	54500	5.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	76306	6.02	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.40%#

Target Compounds					Ovalue
13) Acetone	2.20	43	2648	1.912 ug/L	93
22) cis-1,2-Dichloroethene	3.96	96	163844	13.042 ug/L	98
34) Trichloroethene	5.96	95	11870	0.876 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
Data File : VV010590.D
Acq On : 30 Apr 2019 22:31
Operator : SY/MD
Sample : K2590-18DL 25X
Misc : 25mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XC5DL

Quant Time: May 01 08:57:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 01 08:14:23 2019
Response via : Initial Calibration

