

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010613.D
 Acq On : 01 May 2019 19:01
 Operator : SY/MD
 Sample : K2590-10RE
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0XA2RE

Quant Time: May 02 08:42:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 02 08:18:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	77320	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	77455	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	31735	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22627	4.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	1.57	69	17545	5.99	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.80%
11) 1,1-Dichloroethene-d2	2.13	63	32808	3.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	3.93	46	57720	54.09	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.18%
24) Chloroform-d	4.40	84	46823	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.08	65	26198	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.00%
32) Benzene-d6	5.10	84	99641	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
36) 1,2-Dichloropropane-d6	6.12	67	30612	5.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.00%
41) Toluene-d8	7.36	98	88242	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	7.66	79	9885	4.68	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.60%
46) 2-Hexanone-d5	8.13	63	48566	57.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.18%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	20925	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	32483	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds					Ovalue
9) Trichlorofluoromethane	1.76	101	22043	2.282 ug/L	99
13) Acetone	2.19	43	15421	24.336 ug/L	96
34) Trichloroethene	5.96	95	2998	0.499 ug/L	86
47) Tetrachloroethene	8.02	164	21479	4.528 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050119\
Data File : VV010613.D
Acq On : 01 May 2019 19:01
Operator : SY/MD
Sample : K2590-10RE
Misc : 25mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0XA2RE

Quant Time: May 02 08:42:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 02 08:18:34 2019
Response via : Initial Calibration

