

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010620.D
 Acq On : 01 May 2019 22:23
 Operator : SY/MD
 Sample : VIBLK41
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK41

Quant Time: May 02 08:58:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 02 08:18:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	80857	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	81553	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	31930	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22287	4.69	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.80%
7) Chloroethane-d5	1.58	69	16633	5.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.60%
11) 1,1-Dichloroethene-d2	2.13	63	32042	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	3.96	46	58493	52.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.82%
24) Chloroform-d	4.40	84	48011	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.08	65	25491	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.10	84	98019	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
36) 1,2-Dichloropropane-d6	6.12	67	30655	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	7.36	98	86818	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	7.67	79	9670	4.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.00%
46) 2-Hexanone-d5	8.14	63	45956	51.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.52%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	20999	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	30250	5.04	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050119\
Data File : VV010620.D
Acq On : 01 May 2019 22:23
Operator : SY/MD
Sample : VIBLK41
Misc : 25mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK41

Quant Time: May 02 08:58:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 02 08:18:34 2019
Response via : Initial Calibration

