

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010605.D
 Acq On : 01 May 2019 15:14
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV41

Quant Time: May 02 08:14:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 02 08:11:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	93200	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	92698	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	43623	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	24788	4.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	1.58	69	14202	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.60%
11) 1,1-Dichloroethene-d2	2.13	63	49909	4.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.40%
20) 2-Butanone-d5	3.95	46	62684	48.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.46%
24) Chloroform-d	4.40	84	55893	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.08	65	27057	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
32) Benzene-d6	5.10	84	116117	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.12	67	33664	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	7.36	98	109640	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	7.67	79	12554	4.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.40%
46) 2-Hexanone-d5	8.14	63	50932	50.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.92%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	23722	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	37643	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.14	85	41512	4.890	ug/L 99
3) Chloromethane	1.25	50	33910	4.702	ug/L 96
5) Vinyl chloride	1.32	62	34341	4.715	ug/L 99
6) Bromomethane	1.54	94	19825	4.519	ug/L 96
8) Chloroethane	1.60	64	13058	4.091	ug/L 95
9) Trichlorofluoromethane	1.77	101	58279	5.005	ug/L 96
10) 1,1,2-Trichloro-1,2,2-trif	2.14	101	30154	4.926	ug/L 98
12) 1,1-Dichloroethene	2.14	96	25888	4.655	ug/L 91
13) Acetone	2.21	43	33979	44.486	ug/L 100
14) Carbon disulfide	2.32	76	72404	4.757	ug/L 99
15) Methyl Acetate	2.47	43	11201	4.740	ug/L 97
16) Methylene chloride	2.53	84	31217	4.701	ug/L 96
17) Methyl tert-butyl Ether	2.81	73	63317	4.795	ug/L 98
18) trans-1,2-Dichloroethene	2.79	96	31342	4.920	ug/L 99
19) 1,1-Dichloroethane	3.23	63	57389	4.919	ug/L 98
21) 2-Butanone	4.04	43	67550	52.036	ug/L 99
22) cis-1,2-Dichloroethene	3.96	96	33517	4.901	ug/L 91

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050119\
 Data File : VV010605.D
 Acq On : 01 May 2019 15:14
 Operator : SY/MD
 Sample : VSTDICV005
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV41

Quant Time: May 02 08:14:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR050119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 02 08:11:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
23) Bromochloromethane	4.30	128	14169	4.893	ug/L	99
25) Chloroform	4.43	83	59598	4.780	ug/L	96
27) 1,2-Dichloroethane	5.18	62	34371	4.814	ug/L	95
29) 1,1,1-Trichloroethane	4.66	97	51017	4.792	ug/L	99
30) Cyclohexane	4.72	56	49375	4.928	ug/L	99
31) Carbon tetrachloride	4.87	117	47125	4.947	ug/L	97
33) Benzene	5.15	78	127022	4.973	ug/L	100
34) Trichloroethene	5.96	95	34231	4.759	ug/L	97
35) Methylcyclohexane	6.18	83	51737	4.966	ug/L	97
37) 1,2-Dichloropropane	6.22	63	30701	4.918	ug/L	100
38) Bromodichloromethane	6.55	83	39289	4.936	ug/L	96
39) cis-1,3-Dichloropropene	7.07	75	44145	5.218	ug/L	95
40) 4-Methyl-2-pentanone	7.28	43	171791	50.133	ug/L	98
42) Toluene	7.43	91	140685	4.923	ug/L	99
44) trans-1,3-Dichloropropene	7.69	75	32638	4.946	ug/L	96
45) 1,1,2-Trichloroethane	7.88	97	20890	4.710	ug/L	91
47) Tetrachloroethene	8.02	164	27204	4.792	ug/L	97
48) 2-Hexanone	8.19	43	124681	50.725	ug/L	99
49) Dibromochloromethane	8.29	129	25446	4.927	ug/L	98
50) 1,2-Dibromoethane	8.40	107	19883	5.017	ug/L #	97
51) Chlorobenzene	8.93	112	89582	4.821	ug/L	99
52) Ethylbenzene	9.05	91	150165	4.955	ug/L	99
53) m,p-xylene	9.18	106	58591	5.048	ug/L	99
54) o-xylene	9.59	106	57855	5.119	ug/L	94
55) Styrene	9.60	104	95757	5.107	ug/L	99
56) Isopropylbenzene	9.97	105	150274	5.124	ug/L	99
58) 1,1,1,2,2-Tetrachloroethane	10.29	83	23018	4.595	ug/L	100
59) 1,2,3-Trichloropropane	10.32	75	17475	4.923	ug/L	99
61) Bromoform	9.77	173	12189	4.634	ug/L	98
62) 1,3-Dichlorobenzene	11.23	146	67384	4.814	ug/L	99
63) 1,4-Dichlorobenzene	11.32	146	67730	4.856	ug/L	96
65) 1,2-Dichlorobenzene	11.69	146	64885	4.876	ug/L	100
66) 1,2-Dibromo-3-chloropropan	12.48	75	3587	4.703	ug/L	94
67) 1,3,5-Trichlorobenzene	12.69	180	49852	4.747	ug/L	99
68) 1,2,4-trichlorobenzene	13.31	180	37962	4.818	ug/L	99
69) Naphthalene	13.55	128	59639	4.949	ug/L	99
70) 1,2,3-Trichlorobenzene	13.79	180	36384	5.178	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

