

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV050120\
 Data File : VV015861.D
 Acq On : 01 May 2020 16:29
 Operator : SY/MD
 Sample : L2437-02
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFSE1

Quant Time: May 02 05:12:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 02 00:06:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	121556	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	117603	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	54174	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	32540	5.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	1.58	69	29591	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.12	63	49247	3.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.80%
20) 2-Butanone-d5	3.94	46	97265	47.69	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.38%
24) Chloroform-d	4.38	84	70551	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.06	65	40892	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.08	84	132144	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.09	67	42483	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	7.34	98	117331	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
43) trans-1,3-Dichloropropene-	7.65	79	14057	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.11	63	65149	38.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.04%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	28764	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	42927	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV050120\
Data File : VV015861.D
Acq On : 01 May 2020 16:29
Operator : SY/MD
Sample : L2437-02
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFSE1

Quant Time: May 02 05:12:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR042220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 02 00:06:27 2020
Response via : Initial Calibration

