

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050123\
 Data File : VV030957.D
 Acq On : 01 May 2023 15:51
 Operator : SY/MD
 Sample : 02551-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHE81

Quant Time: May 02 01:34:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 29 01:06:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	128686	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	130323	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	66957	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	41980	3.959	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	79.200%	
7) Chloroethane-d5	1.536	69	36310	4.198	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.000%	
11) 1,1-Dichloroethene-d2	2.063	65	22185	3.954	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.000%	
20) 2-Butanone-d5	3.822	46	85058	45.241	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	90.480%	
24) Chloroform-d	4.262	84	83910	3.982	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	79.600%	
26) 1,2-Dichloroethane-d4	4.953	65	44749	4.177	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
32) Benzene-d6	4.973	84	144591	3.785	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	75.800%	
36) 1,2-Dichloropropane-d6	5.998	67	43264	3.895	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	77.800%	
41) Toluene-d8	7.252	98	130019	3.574	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.400%	
43) trans-1,3-Dichloroprop...	7.564	79	15389	3.696	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.000%	
46) 2-Hexanone-d5	8.037	63	56196	38.521	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.040%	
56) 1,1,2,2-Tetrachloroeth...	10.159	84	33268	4.132	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	82.600%	
66) 1,2-Dichlorobenzene-d4	11.571	152	51082	4.093	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	81.800%	
Target Compounds						
2) Dichlorodifluoromethane	1.108	85	659	0.065	ug/L	94
22) cis-1,2-Dichloroethene	3.828	96	1021	0.122	ug/L #	91
34) Trichloroethene	5.854	95	1850	0.212	ug/L	79
47) Tetrachloroethene	7.915	164	20428	2.450	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050123\
Data File : VV030957.D
Acq On : 01 May 2023 15:51
Operator : SY/MD
Sample : 02551-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHE81

Quant Time: May 02 01:34:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 29 01:06:38 2023
Response via : Initial Calibration

