

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050123\
 Data File : VV030959.D
 Acq On : 01 May 2023 16:37
 Operator : SY/MD
 Sample : 02488-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 02 01:34:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 29 01:06:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	125017	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	126389	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	64354	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	44421	4.313	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.539	69	36029	4.288	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.800%
11) 1,1-Dichloroethene-d2	2.063	65	22949	4.210	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.200%
20) 2-Butanone-d5	3.818	46	80667	44.164	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.320%
24) Chloroform-d	4.259	84	83940	4.101	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.000%
26) 1,2-Dichloroethane-d4	4.956	65	45401	4.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	4.969	84	144953	3.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
36) 1,2-Dichloropropane-d6	5.998	67	42336	3.930	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.600%
41) Toluene-d8	7.252	98	130202	3.691	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.800%
43) trans-1,3-Dichloroprop...	7.564	79	15729	3.895	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.037	63	54914	38.814	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.620%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	32633	4.179	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.600%
66) 1,2-Dichlorobenzene-d4	11.570	152	51887	4.325	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.600%
Target Compounds						Qvalue
16) Methylene chloride	2.455	84	6220	0.654	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050123\
Data File : VV030959.D
Acq On : 01 May 2023 16:37
Operator : SY/MD
Sample : 02488-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 02 01:34:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR042823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 29 01:06:38 2023
Response via : Initial Calibration

