

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050222\
 Data File : VW025770.D
 Acq On : 02 May 2022 16:55
 Operator : SY/MD
 Sample : N2625-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6X5

Quant Time: May 03 05:25:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 03 05:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	141188	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	149699	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62870	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	42661	3.250	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	65.000%	
7) Chloroethane-d5	1.571	69	50647	3.627	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	72.600%	
11) 1,1-Dichloroethene-d2	2.111	63	70023	2.663	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	53.200%#	
20) 2-Butanone-d5	3.896	46	105966	66.366	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	132.740%#	
24) Chloroform-d	4.352	84	103774	5.232	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.600%	
26) 1,2-Dichloroethane-d4	5.037	65	48951	5.728	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.600%	
32) Benzene-d6	5.053	84	187626	4.332	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.600%	
36) 1,2-Dichloropropane-d6	6.072	67	62264	5.155	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.000%	
41) Toluene-d8	7.317	98	150499	3.744	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	74.800%	
43) trans-1,3-Dichloroprop...	7.629	79	12731	3.136	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	62.800%	
46) 2-Hexanone-d5	8.088	63	70066	49.170	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.340%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	41585	5.449	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	54953	5.213	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.200%	
Target Compounds						
25) Chloroform	4.368	83	1879	0.102	ug/L #	72
53) m,p-Xylene	9.146	106	1130	0.058	ug/L	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050222\
Data File : VV025770.D
Acq On : 02 May 2022 16:55
Operator : SY/MD
Sample : N2625-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6X5

Quant Time: May 03 05:25:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 03 05:22:31 2022
Response via : Initial Calibration

