

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW050222\
 Data File : VW025772.D
 Acq On : 02 May 2022 17:41
 Operator : SY/MD
 Sample : N2625-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6X7

Quant Time: May 03 05:26:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 03 05:22:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	144013	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	151523	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	64539	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	42542	3.177	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.600%
7) Chloroethane-d5	1.571	69	48919	3.434	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.600%
11) 1,1-Dichloroethene-d2	2.111	63	68563	2.556	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.200%#
20) 2-Butanone-d5	3.896	46	107768	66.171	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	132.340%#
24) Chloroform-d	4.352	84	100962	4.990	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.034	65	48281	5.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.800%
32) Benzene-d6	5.053	84	185942	4.241	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
36) 1,2-Dichloropropane-d6	6.069	67	60463	4.945	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.317	98	144405	3.549	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.000%
43) trans-1,3-Dichloroprop...	7.629	79	12555	3.055	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.200%
46) 2-Hexanone-d5	8.088	63	72474	50.247	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.500%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42133	5.454	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	54808	5.065	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050222\
Data File : VV025772.D
Acq On : 02 May 2022 17:41
Operator : SY/MD
Sample : N2625-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6X7

Quant Time: May 03 05:26:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 03 05:22:31 2022
Response via : Initial Calibration

