

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
 Data File : VV030975.D
 Acq On : 02 May 2023 11:59
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV231

Quant Time: May 03 06:29:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed May 03 06:25:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4219	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	3870	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2196	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	1686	0.471	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	94.000%		
4) 1,2-Dichloroethane-d4	4.969	65	1482	0.486	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
7) 1,2-Dichloropropane-d6	5.995	67	1303	0.508	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	102.000%		
8) Toluene-d8	7.253	98	4759	0.507	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	102.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	1031	0.513	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	102.000%		
Target Compounds						Qvalue
3) Vinyl chloride	1.280	62	1580	0.485	ug/L	98
6) Trichloroethene	5.852	95	1460	0.479	ug/L	99
9) 1,2-Dibromoethane	8.290	107	872	0.500	ug/L	99
12) 1,2,3-Trichloropropane	10.214	75	713	0.528	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.356	75	167	0.533	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050223\
Data File : VV030975.D
Acq On : 02 May 2023 11:59
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV231

Quant Time: May 03 06:29:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM050223.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed May 03 06:25:35 2023
Response via : Initial Calibration

